

Service order number	4247081	Phone	
Customer's adress	Köbenhavns Universitet Retsmedicinsk Institut 2100 København Ö Frederik V's Vej 11 35326288	Contact person	
		Date call received	17.02.2014
		Date of service	25.03.2014
		PO number	

Equipment number	10309015	Serial number	708000319
Equipment description	INSTRUMENT FREEDOM EVO 150 BASE UNIT		
Symptoms - Group code	090	Symptoms - Coding	030

Solution/Corrective action	Årlig vedligeholdelse iflg Tecan dokumenter "Preventive maintenance" og "Operation Qualification" Instrument testet og funde i orden.
----------------------------	--


BC (Bill Code): Z1 = Good will Z2 = Warranty Z3 = To be invoiced Z4 = Warranty on Repair
 Z5 = Contract Z6 = Installation

Component	Description	Quantity	Serial # Added	Serial # Removed	BC
10619863	CABLE ILID SET 4 PCE. GENESIS FREEDOM	2,00			Z5
10619532	SYRINGE 1.0ML NEW	8,00			Z5
10612501	TIP STD. STA.STEEL PTFE COATED EVO SP	8,00			Z5
10649016	VALVE 3-WAY 120° M6 DILUTER XP SMART	8,00			Z5
10619403	TUBING PIP. FEP 2.5*1.5MM 2800MM SP	8,00			Z5
10619402	TUBE FEP IC M6 3*2MM 650MM SP	8,00			Z5

Instrument status at time of service report

- ☒ Repaired - Normal utilization can continue
☐ Fault/Defect not reproducible - Normal utilization can continue
☐ Fault/Defect still present

Date	Work Hours	BC	Travel Hours	BC	Distance	BC
21.03.2014	8,00	Z5	2,00	Z5		

Date: 25.03.2014	Tecan	Niels Andersen	Customer	No customer
	Signature		Signature	

Purpose

This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

Customer information		Instrument information			
Contact account		Freedom EVO Clinical	100 <input type="checkbox"/>	150 <input type="checkbox"/>	200 <input type="checkbox"/>
Contact person	Brian Schou Rasmussen	Freedom EVO	100 <input type="checkbox"/>	150 <input type="checkbox"/>	200 <input checked="" type="checkbox"/>
Lab site	Retskemisk Afdeling	Freedom EVO 75	<input type="checkbox"/>	8 plus 1 Access	<input type="checkbox"/>
Address	Retsmedicinsk Institut				
	Københavns Universitet	PosID	<input checked="" type="checkbox"/>		
		Arm left	LiHa <input checked="" type="checkbox"/>	AirLiHa <input type="checkbox"/>	RoMa <input type="checkbox"/>
			PnP <input type="checkbox"/>	MCA <input type="checkbox"/>	
	Frederik V's vej 11, 3. Sal	Arm middle	LiHa <input type="checkbox"/>	AirLiHa <input type="checkbox"/>	RoMa <input type="checkbox"/>
			PnP <input type="checkbox"/>	MCA <input type="checkbox"/>	
		Arm right	LiHa <input type="checkbox"/>	AirLiHa <input type="checkbox"/>	RoMa <input checked="" type="checkbox"/>
			PnP <input type="checkbox"/>	MCA <input type="checkbox"/>	
ZIP code + City	2100 København Ø				
Country	DK	LiHa configuration	left <input checked="" type="checkbox"/>	middle <input type="checkbox"/>	right <input type="checkbox"/>
			Te-Fill <input type="checkbox"/>	PMP <input type="checkbox"/>	
Service Engineer	Niels Andersen	No of tips	8	Type of tips	std tip
		No of tips		Type of tips	
		LiHa configuration	left <input type="checkbox"/>	middle <input type="checkbox"/>	right <input type="checkbox"/>
			Te-Fill <input type="checkbox"/>	PMP <input type="checkbox"/>	
		No of tips		Type of tips	
		No of tips		Type of tips	
		MCA configuration	No of tips	Type of tips	
		Serial number of the instrument 708000319			
		Instrument	new installation <input type="checkbox"/>	used before <input checked="" type="checkbox"/>	

Reference documents	Document Number	Document Version	N/A
Freedom EVO Service Manual	392887	V...	<input checked="" type="checkbox"/>
Freedom EVO-2 Service Manual	393828	V.7.4..	<input type="checkbox"/>
Freedom EVO 75 Service Manual	394880	V...	<input checked="" type="checkbox"/>
Freedom EVO Clinical Operating Manual	393062	V...	<input checked="" type="checkbox"/>
Freedom EVO Operating Manual	392886	V.7.3..	<input type="checkbox"/>
Freedom EVO 75 Operating Manual	393248	V...	<input checked="" type="checkbox"/>
Instrument Software Manual	392888	V.5.0..	<input type="checkbox"/>
Proteam Advanced Digest Application Freedom EVO	392513	V...	<input checked="" type="checkbox"/>
Te-PoolSafe IQ/OQ	395255	V...	<input checked="" type="checkbox"/>

Purpose This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

In the following sections place a checkmark in the check box that applies. Many steps within Setup and Service software have the option to print out data/result sheets once the step is complete. When a print option is available: print out, sign, and date the data/result sheets. File the sheets in the print out chapter in this Logbook.

If the operation qualification is being performed immediately after the IQ and the tests listed in these sections passed as part of the IQ – do not repeat them. If the OQ is not being performed immediately after the IQ or for any test(s) that failed in the IQ, repeat the necessary test(s).

Instrument modules	Acceptance values	N/A	Done	Passed
<ul style="list-style-type: none"> • System Devices / Move Test <ul style="list-style-type: none"> - Random Move Test 	Duration 600 cycles, Re-init 200 cycles	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • System Devices / Move Test 2 <ul style="list-style-type: none"> - Random Move Test 	(Use this test in case of an obstructed worktable only, run either the Move Test or the Move Test2) Duration 600 cycles, Re-init 200 cycles	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • System Devices / LiHA <ul style="list-style-type: none"> - Verify Reference Positions - Te-PS (Positioning System) - Tip Adapter Test - Flush tips once sequentially 	Reference positions accurately Tip alignment and individual Z verification passed Test passed with all available DiTi channels Order of pipetting tubing is correct	<input type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input type="checkbox"/>	<input checked="" type="checkbox"/> <input type="checkbox"/> <input checked="" type="checkbox"/>	<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>
<ul style="list-style-type: none"> • System Devices / AirLiHA <ul style="list-style-type: none"> - Verify Reference Positions - Te-PS (Positioning System) - Tip Adapter Test 	Reference positions accurately Tip alignment and individual Z verification passed Test passed with all available DiTi channels	<input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/> <input type="checkbox"/>
<ul style="list-style-type: none"> • Liquid System / DiTi Test <ul style="list-style-type: none"> - (Lower) DiTi Eject Test 	Test passed with all available DiTi channels	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • Liquid System / Te-Fill <ul style="list-style-type: none"> - Channel Sequence - Dispense - Aspiration 	Test passed with all available channels Test passed with all available channels Test passed with all available channels	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>
<ul style="list-style-type: none"> • Liquid System / PMP <ul style="list-style-type: none"> - USB Test - PMP Channel Test - Clogging Test - Leakage Test - P-Sensor Test - Filter Test - Pressure LLD Test 	Test passed Test passed with all available channels Test passed with all available channels Test passed with all available channels Test passed with all available channels Test passed with all available channels Test passed with all available channels	<input checked="" type="checkbox"/> <input type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>
<ul style="list-style-type: none"> • Liquid System / Liquid System <ul style="list-style-type: none"> - Liquid Level Detection Test 	(A minimum of 2 tests need to be performed) Test passed with trough rack Test passed with micro plate Test passed with strip rack	<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>		<input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/>
<ul style="list-style-type: none"> - FaWa Test 	Test passed with all available channels	<input type="checkbox"/>		<input checked="" type="checkbox"/>

4

Purpose

This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

Instrument modules	Acceptance values	N/A	Done	Passed
- Liquid LiHa Gravimetric Precision Test	It's required to perform the tests at all indicated volumes of the appropriate tip configuration with either the Gravimetric Precision Test or QC Kit.			
- LiHa Gravimetric Precision Test - Low Vol tip / Syringe 500µl - Using Low Vol option	Test passed at 1µl CV≤ 10.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- LiHa Gravimetric Precision Test - Low Vol tip / Syringe 500µl	Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- LiHa Gravimetric Precision Test - Te-PS tip / Syringe 250µl	Test passed at 1µl CV≤ 10.0% Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- LiHa Gravimetric Precision Test - Standard tip / Syringe 1000µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 0.75%	<input type="checkbox"/>		<input checked="" type="checkbox"/> <input checked="" type="checkbox"/>
- LiHa Gravimetric Precision Test - Diti Type 10µl / Syringe 500µl - Using Low Vol option	Test passed at 1µl CV≤ 10.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- LiHa Gravimetric Precision Test - Diti Type 10µl / Syringe 500µl	Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- LiHa Gravimetric Precision Test - DiTi 200µl / Syringe 1000µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 0.75%*	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- Gravimetric Precision Test for none-standard configuration	1st measurement = Reference 2nd measurement = Comparison to Reference	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- Liquid LiHa Colorimetric Precision Test with QC Kit	It's required to perform the tests at all indicated volumes of the appropriate tip configuration with either the Gravimetric Precision Test or QC Kit.			
- LiHa QC Kit - Fixed tip Low Vol / Syringe 500µl	Test passed at 1µl CV≤ 10.0% Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- LiHa QC Kit - Fixed tip Low Vol / Syringe 1000µl	Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- LiHa QC Kit - Fixed tip Standard / Syringe 1000µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 1.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- LiHa QC Kit - Diti Type 10µl / Syringe 500µl	Test passed at 1µl CV≤ 10.0% Test passed at 10µl CV≤ 3.5%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- LiHa QC Kit - Diti Type 200µl / Syringe 1000µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 1.0%*	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>

②A
②B

* The different methods mentioned above are intended to verify that the system is operating within its specifications. Please be aware that different measurement methods have different inherent systematic measurement errors and therefore the acceptance values can be different. Details can be found in the operating manual.

Purpose This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

Instrument modules	Acceptance values	N/A	Done	Passed
- Air LiHa Gravimetric Precision Test	It's required to perform the tests at all indicated volumes of the appropriate tip configuration with either the Gravimetric Precision Test or QC Kit.			
- Air LiHa Gravimetric Precision Test Diti Type 200µl / non filtered	Test passed at 100µl CV≤ 0.5% Test passed at 10µl CV≤ 2.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- Air LiHa Gravimetric Precision Test Diti Type 10µl / non filtered	Test passed at 1µl CV≤ 8.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Air LiHa Colorimetric Precision Test with QC Kit	It's required to perform the tests at all indicated volumes of the appropriate tip configuration with either the Gravimetric Precision Test or QC Kit.			
- Air LiHa Colorimetric Precision Test Diti Type 200µl / non filtered	Test passed at 10µl CV≤ 2.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Air LiHa Colorimetric Precision Test Diti Type 10µl / non filtered	Test passed at 1µl CV≤ 8.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- 8 plus 1 Colorimetric Precision Test with S&S software				
- 8 plus 1 Colorimetric Test with Orange G / Sunrise reader only Diti Type 200µl / Syringe 1000 µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 1.75%*	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- 8 plus 1 Colorimetric Precision Test with QC kit				
- 8 plus 1 Colorimetric Precision Test with QC Kit for all channels Infinite series reader only Diti Type 200µl / non filtered Syringe 1000 µl	Test passed at 10µl CV≤ 3.5% Test passed at 100µl CV≤ 1.0%*	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
• System Devices / RoMa		<input checked="" type="checkbox"/>		
- Reference plate test	Passed with 10 cycles			<input type="checkbox"/>
• System Devices / PnP		<input checked="" type="checkbox"/>		
- Verify Reference positions	Reference positions accurately		<input type="checkbox"/>	
- Tube Move Test	Passed at left and right grid position			<input type="checkbox"/>

* The different methods mentioned above are intended to verify that the system is operating within its specifications. Please be aware that different measurement methods have different inherent systematic measurement errors and therefore the acceptance values can be different. Details can be found in the operating manual.

Purpose

This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

• System Devices MCA96		<input checked="" type="checkbox"/>		
- DiTi Test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Tip Block Test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Plate Move Test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA96 Colorimetric Precision Test with QC Kit	It's required to perform the tests at all indicated volumes of the appropriate tip configuration. (3 plates needed to create the pass / fail report)			
- MCA96 QC kit / 50 µl DiTi	Test passed at 1µl CV≤ 6.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA96 QC kit / 100 µl DiTi	Test passed at 1.5µl CV≤ 6.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA96 QC kit / 200 µl DiTi	Test passed at 2µl CV≤ 6.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA96 QC kit / Standard Tipblock	Test passed at 5µl CV≤ 4.0% Test passed at 50µl CV≤ 3.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/>
- MCA96 QC kit / HP Tipblock	Test passed at 1µl CV≤ 10.0% Test passed at 5µl CV≤ 4.0% Test passed at 10µl CV≤ 3.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>
- Tightness Test, performed in EVOware. scripts available in: \\EVOware\database\maintenance	Head is able to aspirate liquid and hold liquid for a minimum time of 5 minutes (Diti's) / 1 minute (FTB), no droplets on the worktable during this time.			<input type="checkbox"/>

Purpose This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

Instrument modules	Acceptance values	N/A	Done	Passed
• System Devices MCA384		<input checked="" type="checkbox"/>		
- Set defaults in EEPROM		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
- Set head production data		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
- Determine range		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
- Determine reference positions	(scaling and offset)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
- Range test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Park sensor test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Plunger init sensor test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Head adapter test	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA384 Colorimetric Precision Test with QC Kit	It's required to perform the tests at all indicated volumes of the appropriate tip configuration. (3 plates needed to create the pass / fail report)			
- MCA384 QC kit / 15µl DiTi	Test passed at 0.5µl CV≤ 4.0% ACC± 10.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA384 QC kit / 50µl DiTi	Test passed at 1.0µl CV≤ 4.0% ACC± 5.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA384 QC kit / 125µl DiTi	Test passed at 2.0µl CV≤ 3.0% ACC± 5.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA384 QC kit / small capacity Tipblock	Test passed at 1.0µl CV≤ 8.0% ACC± 5.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- MCA384 QC kit / large capacity Tipblock	Test passed at 5.0µl CV≤ 4.0% ACC± 5.0%	<input checked="" type="checkbox"/>		<input type="checkbox"/>
- Tightness Test, performed in EVOware. scripts available in: \EVOware\database\maintenance	Head is able to aspirate liquid and hold liquid for a minimum time of 5 minutes, no droplets are falling into the Microplate during this time.			<input type="checkbox"/>

Instrument modules, accessories	Acceptance values	N/A	Done	Passed
• CGM		<input checked="" type="checkbox"/>		
Set defaults in EEPROM		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Calibrate rotator		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Align rotator		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Determine range of y and g axis		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Gripper finger alignment and z-range		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Determine reference positions (scaling an offset)		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Adjust displacement of r and g axis		<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Rotator Encoder Test		<input checked="" type="checkbox"/>		<input type="checkbox"/>
Park sensor test		<input checked="" type="checkbox"/>		<input type="checkbox"/>
Range Test		<input checked="" type="checkbox"/>		<input type="checkbox"/>
Plate move test		<input checked="" type="checkbox"/>		<input type="checkbox"/>

Purpose This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

Instrument modules	Acceptance values	N/A	Done	Passed
<ul style="list-style-type: none"> • System Devices / PosID2 / PosID-3 <ul style="list-style-type: none"> - Reading Position Test (PosID-3) - No Tube Sensor Test - Barcode Reading Test 	Passed Passed Passed with 5 cycles	<input checked="" type="checkbox"/> <input checked="" type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/>
<ul style="list-style-type: none"> • Options / Access Status Options <ul style="list-style-type: none"> - I/O Module tests - Loading interface tests - RSS tests - SPO/MPO sensors tests - Safety tests 	Passed Passed Passed Passed Passed	<input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input type="checkbox"/>		<input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input type="checkbox"/> <input checked="" type="checkbox"/>
<ul style="list-style-type: none"> • Options / Incubator <ul style="list-style-type: none"> - Incubator tests 	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • Options / Te-Link <ul style="list-style-type: none"> - Te-Link tests 	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • Options / Te-MagS <ul style="list-style-type: none"> - Te-MagS tests 	Passed	<input checked="" type="checkbox"/>		<input type="checkbox"/>
<ul style="list-style-type: none"> • Options / Te-Shake <ul style="list-style-type: none"> - Te-Shake tests 	Passed	<input type="checkbox"/>		<input checked="" type="checkbox"/>

③

④

Purpose

This form serves as a checklist for a Tecan authorized Service Technician on the main verification process. However, this form does not contain details (refer to reference documents). It must be filled out after installation and preventive maintenance. This checklist serves for hardware verification only and does not cover any process validation.

• Options / Te-Stack		<input checked="" type="checkbox"/>		
- Te-Stack tests	Passed			<input type="checkbox"/>
• Options / Te-MO		<input checked="" type="checkbox"/>		
- Te-MO tests	Passed			<input type="checkbox"/>
• Options / Te-VacS		<input type="checkbox"/>		
- Te-VacS tests	Passed			<input checked="" type="checkbox"/> (5)
• Centrifuge Rotanta		<input checked="" type="checkbox"/>		
- Imbalance test	No imbalance with 30 g / Imbalance with 45 g			<input type="checkbox"/>
- Temperature sensor test	T1, T2, T3 = room temperature +/- 1° C			<input type="checkbox"/>
- Speed test	RPM 3500 ±3%			<input type="checkbox"/>
• Te-PoolSafe		<input checked="" type="checkbox"/>		
- Te-PoolSafe	Passed according to Doc 395255			<input type="checkbox"/>
Instrument / basic setup	Create EEPROM backup files		<input checked="" type="checkbox"/>	
Instrument / Information	Make a printout		<input checked="" type="checkbox"/>	
Mobile Tool	Maintain equipment characteristics data in SAP/MAM (e.g. SW version)		<input checked="" type="checkbox"/>	

Additional Test	describe test and acceptance values	Done	Passed

The Service Technician confirms that the instrument has been verified in accordance with this operation qualification. Approval by the Lab Manager or responsible person is required.

12mar2014

Niels Andersen, T-NO

Brian Schou Rasmussen

Verification date

Name of Service Technician

Name of Lab Manager

mar 2015



Next due date maintenance/ verification

Signature

Signature

Purpose This form serves as a checklist to enable the Tecan authorized Service Technician a reference guide on the maintenance process, however, this form does not contain details.

Customer information		Instrument information			
Contact account		Freedom EVO Clinical	100 <input type="checkbox"/>	150 <input type="checkbox"/>	200 <input type="checkbox"/>
Contact person	Brian Schou Rasmussen	Freedom EVO 75	<input type="checkbox"/>	100 <input type="checkbox"/>	150 <input type="checkbox"/> 200 <input checked="" type="checkbox"/>
Lab site	Retskemisk Afdeling	Arm left	LiHa <input checked="" type="checkbox"/> AirLiHa <input type="checkbox"/> Te-Fill <input type="checkbox"/> RoMa <input type="checkbox"/> PnP <input type="checkbox"/> MCA <input type="checkbox"/>		
Address	Retsmedicinsk Institut	Arm middle	LiHa <input type="checkbox"/> AirLiHa <input type="checkbox"/> Te-Fill <input type="checkbox"/> RoMa <input type="checkbox"/> PnP <input type="checkbox"/> MCA <input type="checkbox"/>		
	Københavns Universitet	Arm right	LiHa <input type="checkbox"/> AirLiHa <input type="checkbox"/> Te-Fill <input type="checkbox"/> RoMa <input checked="" type="checkbox"/> PnP <input type="checkbox"/> MCA <input type="checkbox"/>		
	Frederik V's vej 11, 3. Sal	LiHa configuration	No of tips 8	Type of tips	std tip
ZIP code + City	2100 København Ø				
Country	DK	Serial number of the instrument 708000319			
Service Engineer	Niels Andersen	Instrument new installation <input type="checkbox"/> used before <input checked="" type="checkbox"/>			

Reference documents	Document Number	Document Version
Freedom EVO Service Manual	392887	V na
Freedom EVO-2 Service Manual	393828	V 7.4
Freedom EVO-75 Service Manual	394880	V na

System	Interval	Variable — depending on stress condition	Action	N/A	Done
Complete Freedom EVO	Every PM*		Decontaminate system		<input checked="" type="checkbox"/>
MCA96 / MCA384	Every PM*		Check tip adapter function	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384	Every PM*		Empty / clean wash / waste containers	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384	Every PM*		Clean wash block	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384	Every PM*		Replace wash liquid filter	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Air filter in dust cover	Every PM*		Replace as needed	<input checked="" type="checkbox"/>	<input type="checkbox"/>
LiHa MultiSense; pipetting tubing	6 months		Cut low volume tubing to remove tubing section worn out by x-ring	<input checked="" type="checkbox"/>	<input type="checkbox"/>
LiHa MultiSense; tip adapter	6 months		Replace Diti Kit MultiSense	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Complete Freedom EVO	12 months		Clean system		<input checked="" type="checkbox"/>
Frontal arm guide	12 months		Clean		<input checked="" type="checkbox"/>
Worktable	12 months		Visually inspection, check grids and replace if necessary		<input checked="" type="checkbox"/>
LiHa	12 months		Visually inspection	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LiHa; support tubing	12 months		Check condition	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LiHa DiTi cone and tubing extension	12 months		Replace	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Air LiHa; DiTi cone tip adapter	12 months		Replace option DiTi cone Air LiHa	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384	12 months		Visually inspection	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA384; s-motor spindle	12 months		Clean and grease s-motor spindle	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384; p-drive	12 months		Check the toothed belt of the p-drive	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384; wash system	12 months		Replace the external tubing	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96 / MCA384; wash system	12 months		Check the internal of the wash station for corrosion and leakage and replace the internal tubing	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96; gripper	12 months		Clean and lubricate gripper spindle	<input checked="" type="checkbox"/>	<input type="checkbox"/>
CGM (MCA384 Gripper)	12 months		Visually inspection	<input checked="" type="checkbox"/>	<input type="checkbox"/>
RoMa	12 months		Visually inspection	<input type="checkbox"/>	<input checked="" type="checkbox"/>
RoMa; Z-rod	12 months		Clean	<input type="checkbox"/>	<input checked="" type="checkbox"/>
PnP	12 months		Visually inspection	<input checked="" type="checkbox"/>	<input type="checkbox"/>

*Every time when maintenance work is performed

System	Interval	Variable — depending on stress condition	Action	N/A	Done
Liquid system, diluters	12 months		Replace syringe	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Liquid system, diluters	12 months		Replace 3-way valve	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LiHa	12 months		Replace fixed tips	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LiHa MultiSense Option	12 months		Replace Multisense tip adapter	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Liquid system	12 months		Replace aspirating tubing	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Liquid system	12 months		Replace interconnecting tubing	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Liquid system	12 months		Replace pipetting tubing	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Liquid system	12 months		Check and replace waste tubing, if necessary.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
LiHa Te-fill option	12 months		Replace complete Te-fill tubing	<input checked="" type="checkbox"/>	<input type="checkbox"/>
LiHa; AirLiHa Z-rod	12 months		Clean and apply a very thin layer of grease	<input checked="" type="checkbox"/>	<input type="checkbox"/>
X-rail	12 months		Clean and apply thin layer of grease		<input checked="" type="checkbox"/>
MCA96; pipetting head	12 months		Replace all 96 tip cone seals	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96; pipetting head	12 months		Lubricate the plungers and spindles	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA384; pipetting head	12 months		Check the gaskets and blunt tubes and replace if necessary	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Centrifuge	12 months		Perform the imbalance test	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Centrifuge	12 months		Carry out the speed calibration	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Centrifuge	12 months		Carry out the temperature calibration	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Centrifuge Rotanta 46	12 months		Check hangers for micro cracks; check the expiry date of the hangers.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
AirLiHa Cylinder Assembly	12 months	After 500'000 plunger moves	Check number of plunger moves according to service manual, Replace Air LiHa cylinder assembly if necessary.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Verifications performed according to Operating Qualification Doc. 392817	12 months		Perform in S&S		<input type="checkbox"/>
Air LiHa; tip adapter	24 months		Replace tip adapter AirLiHa	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Liquid LiHa; tip adapter	24 months		Replace tip adapter on Liquid LiHa used with disposable tips	<input checked="" type="checkbox"/>	<input type="checkbox"/>
LiHa; liquid detection	24 months		Replace ILID cables	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LiHa MultiSense option	24 months		Replace pressure sensor cables	<input checked="" type="checkbox"/>	<input type="checkbox"/>
8 plus 1 ejector blocks	24 months		Replace	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA96; pipetting head	36 months		Replace the pipeting head	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Te-Fill option	36 months		Replace bidirectional pump	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Centrifuge Rotanta 46: Rubber- metal bearings	36 months		Check for cracks; if necessary replace	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MCA384; pipetting head		After 1'000'000 moves	Replace the pipetting head	<input checked="" type="checkbox"/>	<input type="checkbox"/>

The Service Technician confirms that maintenance on the instrument has been performed in accordance with this Preventive Maintenance Checklist.

12mar2014

mar 2015

Niels Andersen, T-NO

Date

next due date maintenance / verification

Service Technician



Liquid Handling System

LiquidSystem.dll Version: 1.10.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 13:07:47

Test Result: Passed
 Tests done: All
 Test Configuration: User Defined

Operator: Service tecan

Date: *12mar2014* Signature: *[Signature]*

Devices

LiHa Arm

Serial Number: 510000060
 Address: C5
 Firmware Version: V1.40-11/2006
 Bootware Version: V1.00-10/2000

Liquid Channel Configuration

	Tip Type	Pipetting Tubing	Syringe Volume [ul]
Tip 1	Fixed Tip Standard	Standard	1000
Tip 2	Fixed Tip Standard	Standard	1000
Tip 3	Fixed Tip Standard	Standard	1000
Tip 4	Fixed Tip Standard	Standard	1000
Tip 5	Fixed Tip Standard	Standard	1000
Tip 6	Fixed Tip Standard	Standard	1000
Tip 7	Fixed Tip Standard	Standard	1000
Tip 8	Fixed Tip Standard	Standard	1000

Aspiration Tubing Configuration

Tubing type: Standard

FaWa

Serial Number: 1001
 Address: C6T30
 Firmware Version: V2.00-11/99
 Bootware Version: V1.00-11/99

Worktable: Retskemisk fixed tip med balance 2014

Balance

Balance type: AG285
 Serial number: 1120503480
 Calibration date: 03okt2013
 Samples [count]: 5
 Tolerance [mg]: 0.10
 Weigh Delay [sec]: 0.5

FaWa Test: Passed

Test Configuration Details

Pump duration before opening the valves[ms]: 1000
 Duration with open valves[ms]: 3000
 Pump duration after closing the valves[ms]: 200
 Tube inner diameter [mm]: 13.0

Pass / Fail Criteria

Minimum expected throughput: see detailed results

Detailed Results

Tips

	Throughput	Min throughput	Within limit
Tip 1, Standard	2030 ul/s	1600 ul/s	Yes
Tip 2, Standard	2039 ul/s	1600 ul/s	Yes
Tip 3, Standard	2083 ul/s	1600 ul/s	Yes
Tip 4, Standard	2101 ul/s	1600 ul/s	Yes
Tip 5, Standard	2052 ul/s	1600 ul/s	Yes
Tip 6, Standard	2021 ul/s	1600 ul/s	Yes
Tip 7, Standard	2044 ul/s	1600 ul/s	Yes
Tip 8, Standard	2061 ul/s	1600 ul/s	Yes

Liquid Level Detection Test: Passed**Test Configuration Details**

LLD

Common

Cycles: 15
 Clot Error Limit [mm]: 4.0
 Error Limit [mm]: 1.5
 Air Gap [ul]: 30
 LLD Speed [mm/sec]: 60.0
 Clot LLD Speed [mm/sec]: 40.0
 Source Liquid Conductivity: Bad
 Prefill LLD Mode: Trough mode
 Prefill Aspiration Speed [ul/sec]: 100
 Prefill Dispense Speed [ul/sec]: 300
 Prefill Submerge [mm]: 2.0
 Retract Speed [mm/sec]: 20.0
 User prompt on error: enabled

LLD Source Racks

	Name	FirstWell	WellCount
Trough (Prefill > 5ml)	QCTROUGH.RF	1	1
Microplate	QCTROUGH.RF	1	1
Strip Rack	QCTROUGH.RF	1	1

LLD Dest Racks

	Name	FirstWell	WellCount	Interleaved Wells
Trough (Prefill > 5ml)	QCTROUGH.RM	1	1	0
Microplate	MP96_RoMa.RF	1	8	0
Strip Rack	QCTUBE1.R1	1	8	0

LLD Prefill Volume

	Tip1 [ul]	Tip2 [ul]	Tip3 [ul]	Tip4 [ul]	Tip5 [ul]	Tip6 [ul]	Tip7 [ul]	Tip8 [ul]
Trough (Prefill > 5ml)	0	0	0	0	0	0	0	0
Microplate	200	200	200	200	200	200	200	200
Strip Rack	600	600	600	600	600	600	600	600

LLD Liquid

	Submerge [mm]	Liquid Conductivity	LLD Mode
Trough (Prefill > 5ml)	2.0	Good	Trough mode
Microplate	2.0	Bad	Odd / even tips twice
Strip Rack	2.0	Good	Odd / even tips twice

Pass / Fail Criteria

'z-in-dev' smaller or equal 'Error Limit [mm]': 1.5

'Liquid det err' equals 0

'Clot error' equals 0

How these results are achieved:

For tips that are expected to find liquid (prefill volume > 0 or destination is a trough):

- 1) 'Liquid det err' is incremented if no liquid is detected.
- 2) 'z-in-dev': max difference of the found levels measured over 'Cycles'.
- 3) 'Clot error' is incremented if no exit signal occurs within 'Clot Error Limit'.

For tips that are not expected to find liquid (no prefill and destination is not a trough):

- 1) 'Liquid det err' is incremented if liquid is detected.
- 2) 'Clot error' is incremented if an exit signal occurs within 'Clot Error Limit'.

Detailed Results

Summary

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-dev	0.3	0.1	0.2	0.2	0.1	0.2	0.2	0.2
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Trough (Prefill > 5ml)

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	42.9	43.1	42.8	42.9	43.1	42.9	43.0	43.5
z-in-max	42.9	43.2	43.0	43.0	43.2	43.0	43.0	43.6
z-in-dev	0.0	0.1	0.2	0.1	0.1	0.1	0.0	0.1
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Microplate

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	72.4	72.1	71.7	71.8	72.0	71.6	71.7	72.0
z-in-max	72.7	72.2	71.8	71.9	72.0	71.8	71.8	72.2
z-in-dev	0.3	0.1	0.1	0.1	0.0	0.2	0.1	0.2
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Strip Rack

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	16.8	17.2	17.0	16.5	17.0	17.0	16.6	17.2
z-in-max	16.8	17.2	17.0	16.7	17.0	17.2	16.8	17.2
z-in-dev	0.0	0.0	0.0	0.2	0.0	0.2	0.2	0.0
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Gravimetric Pipetting Precision Test: Passed

Test Configuration Details

Common

Use next tip for each dispense: enabled
 Liquid type: Tap water
 Density [mg/ul]: 1.000
 Liquid Conductivity: Bad
 LLD Mode: Trough mode

Fixed Tip

Standard

Pipetting Tubing: Standard

10 ul

Test

Cycles: 12

Max CV [%]: 3.500

Aspiration

Aliquotes: 1
Aspiration Delay [msec]: 200
Aspiration Retract Speed [mm/sec]: 20.0
Aspiration Speed [ul/sec]: 30
Aspiration Submerge [mm]: 1.0
Aspiration Volume [ul]: 10.000
Calibration Volume [ul]: 1.000
Detect Speed [mm/sec]: 60
Low-Volume by Aspiration: disabled

Dispense

Break Off Speed [ul/sec]: 150
Delay Before Pinch [msec]: 500
Dispense Delay [sec]: 0.0
Dispense Speed [ul/sec]: 600
Dispense Volume [ul]: 10.000
Dispense by Dilutor: enabled
Dispense on LL: disabled
Dispense on LL Offset [mm]: 0.0
Low-Volume Active: disabled

Liquid Structure

Air Gap Aspiration Speed [ul/sec]: 70
Conditioning Volume [ul]: 0.000
Conditioning Volumes count: 0
Delay After Conditioning [msec]: 0
Excess Volume [ul]: 0.000
Leading Air Gap [ul]: 10.000
Partition Leading Air Gap [ul]: 0.000
Partition Trailing Air Gap [ul]: 0.000
Partition Volume [ul]: 0.000
System Trailing Air Gap [ul]: 20.000
Trailing Air Gap [ul]: 5.000

Wash

Low-Volume by Wash: disabled
Wash Speed [ul/sec]: 1000
Wash Volume [ul]: 1000

Pass / Fail Criteria

CV of each channel and over all CV: less or equal 'Max CV [%]'.

No '0' dispenses are allowed. Limit for '0' dispenses: 10% of nominal volume.

History

	Channels	Start time	Operator comment
1. Validation	1, 2, 3, 4, 5, 6, 7, 8	12/Mar/2014 12:18:22	First run

Detailed Results

Fixed Tip / Standard / Pipetting Tubing: Standard / 10 ul

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8	All
Meas 1	9.780	10.020	10.080	10.140	9.920	10.100	9.820	10.260	-
Meas 2	10.080	10.200	10.000	10.400	9.900	9.980	9.780	9.800	-
Meas 3	10.100	9.860	9.960	10.100	10.100	10.180	9.800	9.560	-
Meas 4	9.960	10.060	10.000	10.040	10.100	10.040	9.900	10.000	-
Meas 5	10.080	10.080	10.080	10.260	9.960	10.240	9.900	9.940	-
Meas 6	10.120	9.940	9.820	10.040	10.080	10.020	9.900	9.680	-
Meas 7	10.000	9.800	10.100	10.280	10.080	9.880	9.960	9.900	-
Meas 8	10.100	10.000	10.080	10.400	10.000	9.940	10.040	9.640	-
Meas 9	10.280	10.060	10.000	9.680	9.800	10.000	9.800	9.640	-
Meas 10	9.860	10.120	9.900	9.760	9.780	9.880	10.160	9.380	-
Meas 11	10.180	10.080	10.040	9.900	9.780	9.920	10.060	9.820	-
Meas 12	10.140	10.140	10.080	9.600	9.680	10.000	9.980	9.940	-
Mean [mg]	10.057	10.030	10.012	10.050	9.932	10.015	9.925	9.797	9.977
Mean [ul]	10.057	10.030	10.012	10.050	9.932	10.015	9.925	9.797	9.977
Acc [%]	0.567	0.300	0.117	0.500	-0.683	0.150	-0.750	-2.033	-0.229
CV [%]	1.370	1.152	0.850	2.681	1.468	1.120	1.200	2.385	1.786

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8	All
Max CV [%]	3.500	3.500	3.500	3.500	3.500	3.500	3.500	3.500	3.500
Status	passed	passed	passed	passed	passed	passed	passed	passed	passed

Colorimetric Pipetting Precision Test: Not Applicable

Test Configuration Details

Pass / Fail Criteria

Detailed Results

Liquid Handling System

LiquidSystem.dll Version: 1.10.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 14:05:02


Test Result: Passed
 Tests done: All
 Test Configuration: User Defined

Operator: Service tecan

Date:

12mar 2014

Signature:



Devices

LiHa Arm

Serial Number: 510000060
 Address: C5
 Firmware Version: V1.40-11/2006
 Bootware Version: V1.00-10/2000

Liquid Channel Configuration

	Tip Type	Pipetting Tubing	Syringe Volume [ul]
Tip 1	Fixed Tip Standard	Standard	1000
Tip 2	Fixed Tip Standard	Standard	1000
Tip 3	Fixed Tip Standard	Standard	1000
Tip 4	Fixed Tip Standard	Standard	1000
Tip 5	Fixed Tip Standard	Standard	1000
Tip 6	Fixed Tip Standard	Standard	1000
Tip 7	Fixed Tip Standard	Standard	1000
Tip 8	Fixed Tip Standard	Standard	1000

Aspiration Tubing Configuration

Tubing type: Standard

FaWa

Serial Number: 1001
 Address: C6T30
 Firmware Version: V2.00-11/99
 Bootware Version: V1.00-11/99

Worktable: Retskemisk fixed tip med balance 2014 Balance

Balance type: AG285
 Serial number: 1120503480
 Calibration date: 03okt2013
 Samples [count]: 5
 Tolerance [mg]: 0.10
 Weigh Delay [sec]: 0.5

FaWa Test: Passed

Test Configuration Details

Pump duration before opening the valves[ms]: 1000
 Duration with open valves[ms]: 3000
 Pump duration after closing the valves[ms]: 200
 Tube inner diameter [mm]: 13.0

Pass / Fail Criteria

Minimum expected throughput: see detailed results

Detailed Results

Tips

	Throughput	Min throughput	Within limit
Tip 1, Standard	2030 ul/s	1600 ul/s	Yes
Tip 2, Standard	2039 ul/s	1600 ul/s	Yes
Tip 3, Standard	2083 ul/s	1600 ul/s	Yes
Tip 4, Standard	2101 ul/s	1600 ul/s	Yes
Tip 5, Standard	2052 ul/s	1600 ul/s	Yes
Tip 6, Standard	2021 ul/s	1600 ul/s	Yes
Tip 7, Standard	2044 ul/s	1600 ul/s	Yes
Tip 8, Standard	2061 ul/s	1600 ul/s	Yes

Liquid Level Detection Test: Passed**Test Configuration Details**

LLD

Common

Cycles: 15
 Clot Error Limit [mm]: 4.0
 Error Limit [mm]: 1.5
 Air Gap [ul]: 30
 LLD Speed [mm/sec]: 60.0
 Clot LLD Speed [mm/sec]: 40.0
 Source Liquid Conductivity: Bad
 Prefill LLD Mode: Trough mode
 Prefill Aspiration Speed [ul/sec]: 100
 Prefill Dispense Speed [ul/sec]: 300
 Prefill Submerge [mm]: 2.0
 Retract Speed [mm/sec]: 20.0
 User prompt on error: enabled

LLD Source Racks

	Name	FirstWell	WellCount
Trough (Prefill > 5ml)	QCTROUGH.RF	1	1
Microplate	QCTROUGH.RF	1	1
Strip Rack	QCTROUGH.RF	1	1

LLD Dest Racks

	Name	FirstWell	WellCount	Interleaved Wells
Trough (Prefill > 5ml)	QCTROUGH.RM	1	1	0
Microplate	MP96_RoMa.RF	1	8	0
Strip Rack	QCTUBE1.R1	1	8	0

LLD Prefill Volume

	Tip1 [ul]	Tip2 [ul]	Tip3 [ul]	Tip4 [ul]	Tip5 [ul]	Tip6 [ul]	Tip7 [ul]	Tip8 [ul]
Trough (Prefill > 5ml)	0	0	0	0	0	0	0	0
Microplate	200	200	200	200	200	200	200	200
Strip Rack	600	600	600	600	600	600	600	600

LLD Liquid

	Submerge [mm]	Liquid Conductivity	LLD Mode
Trough (Prefill > 5ml)	2.0	Good	Trough mode
Microplate	2.0	Bad	Odd / even tips twice
Strip Rack	2.0	Good	Odd / even tips twice

Pass / Fail Criteria

'z-in-dev' smaller or equal 'Error Limit [mm]': 1.5

'Liquid det err' equals 0

'Clot error' equals 0

How these results are achieved:

For tips that are expected to find liquid (prefill volume > 0 or destination is a trough):

- 1) 'Liquid det err' is incremented if no liquid is detected.
- 2) 'z-in-dev': max difference of the found levels measured over 'Cycles'.
- 3) 'Clot error' is incremented if no exit signal occurs within 'Clot Error Limit'.

For tips that are not expected to find liquid (no prefill and destination is not a trough):

- 1) 'Liquid det err' is incremented if liquid is detected.
- 2) 'Clot error' is incremented if an exit signal occurs within 'Clot Error Limit'.

Detailed Results

Summary

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-dev	0.3	0.1	0.2	0.2	0.1	0.2	0.2	0.2
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Trough (Prefill > 5ml)

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	42.9	43.1	42.8	42.9	43.1	42.9	43.0	43.5
z-in-max	42.9	43.2	43.0	43.0	43.2	43.0	43.0	43.6
z-in-dev	0.0	0.1	0.2	0.1	0.1	0.1	0.0	0.1
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Microplate

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	72.4	72.1	71.7	71.8	72.0	71.6	71.7	72.0
z-in-max	72.7	72.2	71.8	71.9	72.0	71.8	71.8	72.2
z-in-dev	0.3	0.1	0.1	0.1	0.0	0.2	0.1	0.2
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Strip Rack

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	16.8	17.2	17.0	16.5	17.0	17.0	16.6	17.2
z-in-max	16.8	17.2	17.0	16.7	17.0	17.2	16.8	17.2
z-in-dev	0.0	0.0	0.0	0.2	0.0	0.2	0.2	0.0
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Gravimetric Pipetting Precision Test: Passed

Test Configuration Details

Common

Use next tip for each dispense: enabled
 Liquid type: Tap water
 Density [mg/ul]: 1.000
 Liquid Conductivity: Bad
 LLD Mode: Trough mode

Fixed Tip

Standard

Pipetting Tubing: Standard
 100 ul

Test

Cycles: 12

Max CV [%]: 0.750

Aspiration

Aliquotes: 1
Aspiration Delay [msec]: 200
Aspiration Retract Speed [mm/sec]: 20.0
Aspiration Speed [ul/sec]: 150
Aspiration Submerge [mm]: 2.0
Aspiration Volume [ul]: 100.000
Calibration Volume [ul]: 3.000
Detect Speed [mm/sec]: 60
Low-Volume by Aspiration: disabled

Dispense

Break Off Speed [ul/sec]: 150
Delay Before Pinch [msec]: 500
Dispense Delay [sec]: 0.0
Dispense Speed [ul/sec]: 600
Dispense Volume [ul]: 100.000
Dispense by Dilutor: enabled
Dispense on LL: disabled
Dispense on LL Offset [mm]: 0.0
Low-Volume Active: disabled

Liquid Structure

Air Gap Aspiration Speed [ul/sec]: 70
Conditioning Volume [ul]: 0.000
Conditioning Volumes count: 0
Delay After Conditioning [msec]: 0
Excess Volume [ul]: 0.000
Leading Air Gap [ul]: 0.000
Partition Leading Air Gap [ul]: 0.000
Partition Trailing Air Gap [ul]: 0.000
Partition Volume [ul]: 0.000
System Trailing Air Gap [ul]: 20.000
Trailing Air Gap [ul]: 10.000

Wash

Low-Volume by Wash: disabled
Wash Speed [ul/sec]: 1000
Wash Volume [ul]: 1000

Pass / Fail Criteria

CV of each channel and over all CV: less or equal 'Max CV [%]'.

No '0' dispenses are allowed. Limit for '0' dispenses: 10% of nominal volume.

History

	Channels	Start time	Operator comment
1. Validation	1, 2, 3, 4, 5, 6, 7, 8	12/Mar/2014 13:14:22	First run

Detailed Results

Fixed Tip / Standard / Pipetting Tubing: Standard / 100 ul

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8	All
Meas 1	98.760	99.400	99.000	99.500	98.960	99.380	98.960	98.720	-
Meas 2	99.120	99.000	98.920	99.300	98.920	99.580	98.700	98.680	-
Meas 3	99.300	99.220	98.820	99.160	99.020	99.500	99.000	98.560	-
Meas 4	99.100	99.140	99.020	99.100	99.020	99.440	98.780	98.840	-
Meas 5	99.000	99.100	98.940	99.000	99.020	99.680	98.760	98.680	-
Meas 6	99.040	99.240	99.100	99.080	98.940	99.720	98.720	98.620	-
Meas 7	99.100	99.080	99.140	99.140	98.800	99.580	98.860	98.580	-
Meas 8	98.820	99.160	99.200	99.020	98.860	99.520	98.720	98.460	-
Meas 9	98.920	99.200	99.000	99.420	98.880	99.600	98.840	98.840	-
Meas 10	98.900	99.140	98.940	99.640	98.860	99.560	98.640	98.780	-
Meas 11	99.380	99.680	98.980	99.700	99.000	99.660	99.140	98.720	-
Meas 12	99.000	99.020	99.180	99.380	98.880	99.620	98.820	98.700	-
Mean [mg]	99.037	99.198	99.020	99.287	98.930	99.570	98.828	98.682	99.069
Mean [ul]	99.037	99.198	99.020	99.287	98.930	99.570	98.828	98.682	99.069
Acc [%]	-0.963	-0.802	-0.980	-0.713	-1.070	-0.430	-1.172	-1.318	-0.931
CV [%]	0.183	0.187	0.116	0.242	0.076	0.099	0.146	0.115	0.304

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8	All
Max CV [%]	0.750	0.750	0.750	0.750	0.750	0.750	0.750	0.750	0.750
Status	passed	passed	passed	passed	passed	passed	passed	passed	passed

Colorimetric Pipetting Precision Test: Not Applicable

Test Configuration Details

Pass / Fail Criteria

Detailed Results

Safety Device Test

Safety.dll Version: 1.13.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Safety Serial Number: 7230367
 Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 14:07:37

Test Result: Passed
 Tests done: All
 Test Configuration: User Defined
 Device Default Settings: na

Operator: Service tecan

Date:

12 Mar 2014

Signature:

[Signature]

Device

Configuration

Firmware Version: V1.30-04/2008
 Bootware Version: V1.10-12/99

Available Options

Door Lock 1 (left)
 Door Lock 2 (right)
 Alarm Device: standard
 Pause / Resume Button

Door Lock Test: Passed

Test Configuration Details

The door lock test is performed once

Pass / Fail Criteria

All questions about the door locks are confirmed with OK
 The sensors recognize the open/locked status correctly

Detailed Results

Door Lock 1 (left): Passed
 Door Lock 2 (right): Not configured for test

Pause / Resume Button Test: Passed

Test Configuration Details

The Pause / Resume button test is performed once

Pass / Fail Criteria

Pause Button interrupts received correctly

Detailed Results

Pause / Resume Button: Passed

Alarm Device Test: Passed

Test Configuration Details

The alarm device test is performed once

Pass / Fail Criteria

All questions about the alarm device are confirmed with OK

Detailed Results

Alarm Green: Passed

Alarm Red / Acoustic: Passed

Te-Shake Device Test

TeShake.dll Version: 1.17.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Te-Shake Serial Number: 7990
 Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 14:09:15

Test Result: Passed
Tests done: Not All
Test Configuration: Default
Device Default Settings: Default

Operator: Service tecan

Date:

12 Mar 2014

Signature:

Ne

Device

Configuration

Shaker Type: 1 Heating Plate
 Firmware Version: V1.10-07/2001
 Bootware Version: V1.20-09/99

Diagnostics

Power up Counter: 857
 Operating Time [minutes]: 281436
 Initialization Counter: 2350
 Movement Counter: 9506
 Overload Counter: 0

Plates

Eppendorf [Plate Type]: 1

Mechanical Burn-In / Validation: Passed

Test Configuration Details

Shake Time [sec]: 60
 Minimum Speed [rpm]: 1200
 Maximum Speed [rpm]: 1200
 Interval Time [sec]: 30
 Interval Step [rpm]: 100
 Heating Block Connected: Yes

Pass / Fail Criteria

Maximum Number of Initialization Errors: 0
 Maximum Move errors: 0

Detailed Results

Move Errors: 0
 Initialization Errors: 0

Test Configuration: Default

Heating Burn-In: Not Done

Test Configuration Details

Pass / Fail Criteria

Detailed Results

Heating Validation with TEMPO110: Not Done

Optional test to validate the heating plate surface temperature on request of customer (on site)

Test Configuration Details

Pass / Fail Criteria

Detailed Results

Te-VacS Device Test

TeVacS.dll Version: 1.15.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Te-VacS Serial Number: 7019
 Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 14:14:01

Test Result: Passed
 Tests done: All
 Test Configuration: User Defined
 Device Default Settings: na

Operator: Service tecan

Date:

12 Mar 2014

Signature:

[Signature]

Liquid Level Sensor Test: Passed
Pressure Sensor Calibration Test: Passed
Extraction Valve Function Test: Passed
Pump Performance Test: Passed
Ventilation Valve Function Test: Passed
Leakage Test: Passed

Device Settings

Type: Chemical resistant

Duty Cycle timing

On Time [ms]: 6
 Off Time [ms]: 16

Duty Cycle usage

Control Valve: off
 Extraction Valve 1: off
 Ventilation Valve 1: off
 Extraction Valve 2: off
 Ventilation Valve 2: off

Test Configuration

External gauge used: no
 Vacuum blocks used: 1
 Ambient pressure equalisation timeout [s]: 15
 Pressure and Timeout

	Pressure [kPa]	Timeout [s]
Maximum pump performance	70	50
Extraction valve	45	3
Leakage test initial	60	40
Maximum leakage	3	60
Ventilation valve	12	3

Detailed Results

Ambient Pressure Equalisation Time [s]: 4.313

Extraction Valve

	Maximum Pump Performance Time [s]	Flowrate Pressure [kPa]
Valve 1	11.547	65
Valve 2	11.016	66

Ventilation Valve

	Leakage Test Initial Time [s]	Leakage Pressure [kPa]	Flowrate Pressure [kPa]
Valve 1	9.891	2	31
Valve 2	9.813	3	30

Instrument Information

Information.dll Version: 1.15.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 09:20:48

Operator: Service tecan

Date:

Signature:

Instrument Information

Instrument Properties

Type: freedom evo
 Size: 150
 Arm order

	Arm	Addr	Movable Range [mm]	Deviation [mm]	Deviation Limit [mm]
1	LiHa	C5	1002.3	1.3	na
2	RoMa	C1	1002.6	1.6	na

Diagnostics Data

Power Ups: 1039
 On Time [h]: 18988.57
 Downloads: 4
 Page erases: 129

Device Information

Device	Firmware	Bootware	Serial Number
C2T00	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-

Device	Firmware	Bootware	Serial Number
C2T02	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C2T04	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T00	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T02	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T04	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T06	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T08	DCSERVO2 V1.21-04/2007	V1.00-05/2003	-
C6T0A	DCSERVO2 V1.20-08/2005	V1.00-05/2003	-
C6T20	XP2000 V1.10-03/2006	V1.02-04/2002	504217493
C6T21	XP2000 V1.10-03/2006	V1.02-04/2002	706003100
C6T22	XP2000 V1.10-03/2006	V1.02-04/2002	706003101
C6T23	XP2000 V1.10-03/2006	V1.02-04/2002	706003102
C6T24	XP2000 V1.10-03/2006	V1.02-04/2002	706003103
C6T25	XP2000 V1.10-03/2006	V1.02-04/2002	706007314
C6T26	XP2000 V1.10-03/2006	V1.02-04/2002	706007315
C6T27	XP2000 V1.10-03/2006	V1.02-04/2002	706007316
C5	LIHACU V1.40-11/2006	V1.00-10/2000	510000060
C6T30	MPO V2.00-11/99	V1.00-11/99	1001
O7	ORBI V1.10-07/2001	V1.20-09/99	7990
C1	ROMACU V2.21-09/2007	V1.00-10/2000	708000317
O2	SAFY V1.30-04/2008	V1.10-12/99	7230367
O5	SPE V2.10-12/01	V1.00-11/99	7019
M1	TECU V1.40-12/2007	V1.10-07/2005	708000319

LiHa Arm

Address: C5
Serial Number: 510000060
Firmware Version: V1.40-11/2006
Bootware Version: V1.00-10/2000
Tips: 8
Spacing: 9.0 - 38.0 mm variable
Lower DiTi Eject: available
LLD Type: Standard ilid

Axes Parameters						
Offset	Displ	Range	Scale	Accel	Speed	Move Speed
X	1.0	-15.1	987.2	0.9996	160.0	1000.0
Y	1.0	-88.9	284.5	1.0000	240.0	350.0
Ys	1.0	9.0	38.0	1.0012	240.0	350.0
Z1	8.5	39.0	171.0	1.0000	200.0	400.0
Z2	8.5	39.0	171.0	1.0000	200.0	400.0
Z3	8.5	39.0	171.0	1.0000	200.0	400.0
Z4	8.4	39.0	171.0	1.0000	200.0	400.0
Z5	8.5	39.0	171.0	1.0000	200.0	400.0
Z6	8.4	39.0	171.0	1.0000	200.0	400.0

	Offset	Displ	Range	Scale	Accel	Speed	Move Speed
Z7	8.5	39.0	171.0	1.0000	200.0	400.0	40.0
Z8	8.5	39.0	171.0	1.0000	200.0	400.0	40.0

Diagnostics Data

	Moves	Moves (cl)	Distance	No-Loads	Fetches	DiTi's	Piercings	Status
X	122893	122893	24470	24	-	-	-	ready
Y	112041	112041	9194	15	-	-	-	ready
Ys	112048	112048	9158	9	-	-	-	ready
Z1	418760	418760	23821	94	11445	0	0	ready
Z2	358038	358038	21832	50	10231	0	0	ready
Z3	352405	352405	21417	100	9791	0	0	ready
Z4	351959	351959	21406	65	9797	0	0	ready
Z5	349130	349130	21229	52	9649	0	0	ready
Z6	348713	348713	21230	51	9648	0	0	ready
Z7	347511	347511	21142	57	9612	0	0	ready
Z8	339770	339770	21594	60	9733	0	0	ready

Te-Fill: NA

RoMa Arm

Address: C1
Serial Number: 708000317
Firmware Version: V2.21-09/2007
Bootware Version: V1.00-10/2000
Type: Eccentric Gripper, CW Rotator, Standard Z-Range

Axes Parameters

	Offset	Displ	Range	Scale	Accel	Speed	Move Speed
X	2.0	127.9	1130.5	0.9997	160.0	1000.0	100.0
Y	1.0	-61.0	323.8	1.0180	100.0	350.0	80.0
Z	1.0	0.0	253.9	1.0000	85.0	130.0	50.0
R	4.4	0.0	275.5	0.9950	60.0	200.0	60.0
G	2.0	56.0	139.0	1.0000	30.0	40.0	10.0

Gripper PWM [%]: 75

Diagnostics Data

Gripped Plates: 9300

Axes

	Moves	Moves (cl)	Distance	No-Loads	Status
X	44910	44910	6723	10	ready
Y	46153	46153	4036	1	ready
Z	135848	135848	5625	1	ready
R	27227	27227	2356	13	ready
G	22571	22571	513	24	ready

Version Information

System Modules

Version	Description	Copyright	Original Filename	Product Name
6.3.13.0	Instrument Setup and Service Application	Copyright © Tecan Schweiz AG 2004 - 2009	SnSFrame.exe	Setup and Service
1.17.13.401	Device library for GENESIS Instruments	Copyright © Tecan Schweiz AG 2008	Genesis.dll	Setup and Service
1.17.13.0	Base classes for concrete Panel Modules	Copyright © Tecan Schweiz AG 2008	Panel.DLL	Setup and Service
1.15.13.0	C++ wrapper for WinApp functions	Copyright © Tecan Schweiz AG 2008	OSpp.dll	Setup and Service
1.16.13.0	GUI Extensions based on MFC controls. Used by Panel base classes and concrete panels	Copyright © Tecan Schweiz AG 2008	GUIExtensions.dll	Setup and Service
1.2.13.0	TCSDriver Tecan Communication Server	Copyright © Tecan Schweiz AG 2008	TCSDriver.dll	Setup and Service
1.5.13.0	TLSDriver - Tecan Login Server	Copyright © Tecan Schweiz AG 2008	TLSDriver.dll	Setup and Service
1.1.3.1	zlib data compression library	(C) 1995-1998 Jean-loup Gailly & Mark Adler	zlib.dll	Setup and Service ZLib.DLL
3.60.0.4	Virtual Print Engine Professional Edition	Copyright © 1995 - 2005 IDEAL Software® GmbH. All rights reserved.	-	Virtual Print Engine Professional Edition

Panels

Version	Description	Copyright	Original Filename	Product Name	Product Version
1.17.13.0	Setup module for basic setups	Copyright © Tecan Schweiz AG 2008	BasicSetup.dll	Setup and Service	6.3
1.8.13.0	Tool to check carrier positions according to worktable DB	Copyright © Tecan Schweiz AG 2008	CheckCarrierPosition.DLL	Setup and Service	6.3
1.15.13.0	Tool module for FW Commands	Copyright © Tecan Schweiz AG 2008	CommandTool.dll	Setup and Service	6.3
1.10.13.0	Setup and QC Test module for DiTiTest	Copyright © Tecan Schweiz AG 2008	DiTiTest.DLL	Setup and Service	6.3
2.3.13.0	Setup and QC Test module for Incubator	Copyright © Tecan Schweiz AG 2008	Incubator.DLL	Setup and Service	6.3
1.15.13.0	Tool module for Information	Copyright © Tecan Schweiz AG 2008	Information.dll	Setup and Service	6.3
1.13.13.0	Setup and QC Test module for I/O-Option	Copyright © Tecan Schweiz AG 2008	iomodule.DLL	Setup and Service	6.3
1.17.13.0	Setup and QC Test module for LiHa	Copyright © Tecan Schweiz AG 2008	LiHa.dll	Setup and Service	6.3
1.10.13.0	Setup and QC Test module for Liquid System	Copyright © Tecan Schweiz AG 2008	Liquid System.DLL	Setup and Service	6.3
1.13.13.0	Setup and QC Test module for Loading Interface of EVO	Copyright © Tecan Schweiz AG 2008	LoadingInterface.DLL	Setup and Service	6.3
1.17.13.0	QC Test Module for RoMa, LiHa and PosID Move Tests	Copyright © Tecan Schweiz AG 2008	MoveTest.dll	Setup and Service	6.3

	Version	Description	Copyright	Original Filename	Product Name	Product Version
PMP.dll	1.10.13.0	Setup and QC Test module for PMP	Copyright © Tecan Schweiz AG 2008	PMP.DLL	Setup and Service	6.3
PosID2.dll	1.14.13.0	Setup and QC Test module for PosID2	Copyright © Tecan Schweiz AG 2008	PosID2.DLL	Setup and Service	6.3
PosID3.dll	1.10.13.0	Setup and QC Test module for PosID3	Copyright © Tecan Schweiz AG 2008	PosID3.DLL	Setup and Service	6.3
Repositioner.dll	1.14.13.0	Setup and QC Test module for Repositioner	Copyright © Tecan Schweiz AG 2008	Repositioner.DLL	Setup and Service	6.3
Results.dll	1.13.13.0	Setup and QC Test module for Results	Copyright © Tecan Schweiz AG 2008	Results.DLL	Setup and Service	6.3
RoboticDevices.dll	1.10.13.0	Tool module for robotic devices	Copyright © Tecan Schweiz AG 2008	RoboticDevices.DLL	Setup and Service	6.3
RoMa.dll	1.17.13.0	Setup and QC Test module for RoMa	Copyright © Tecan Schweiz AG 2008	RomaModule.DLL	Setup and Service	6.3
Safety.dll	1.13.13.0	Setup and QC Test module for Safety	Copyright © Tecan Schweiz AG 2008	Safety.DLL	Setup and Service	6.3
SpoMpo.dll	1.13.13.0	Setup and QC Test module for SPO-MPO	Copyright © Tecan Schweiz AG 2008	SpoMpo.DLL	Setup and Service	6.3
Supervisor2.dll	1.13.13.0	Setup and QC Test module for Supervisor2	Copyright © Tecan Schweiz AG 2008	Supervisor2.DLL	Setup and Service	6.3
TeFill.dll	1.6.13.0	Setup and QC Test module for Te-Fill	Copyright © Tecan Schweiz AG 2008	TeFill.DLL	Setup and Service	6.3
TeLink.dll	1.14.13.0	Setup and QC Test module for Shuttle	Copyright © Tecan Schweiz AG 2008	Shuttle.DLL	Setup and Service	6.3
TeMags.dll	1.17.13.0	Setup and QC Test module for Te-MagS	Copyright © Tecan Schweiz AG 2008	TeMags.dll	Setup and Service	6.3
TeShake.dll	1.17.13.0	Setup and QC Test module for Te-Shake	Copyright © Tecan Schweiz AG 2008	TeShake.DLL	Setup and Service	6.3
TeVacs.dll	1.15.13.0	Setup and QC Test module for Te-VacS	Copyright © Tecan Schweiz AG 2008	TeVacs.dll	Setup and Service	6.3

Liquid Handling System

LiquidSystem.dll Version: 1.10.13.0
 Panel.dll Version: 1.17.13.0
 Genesis.dll Version: 1.17.13.401
 GUIExtensions.dll Version: 1.16.13.0
 OSpp.dll Version: 1.15.13.0
 Setup and Service Version: 6.3.13.0
 Computer Name: PROD-2655.prodnet.internal

Instrument Type: EVO
 Instrument Serial Number: 708000319
 Tools [Type, SN]: None specified
 Date: 12/Mar/2014 12:04:37

Test Result: Passed
 Tests done: Not All
 Test Configuration: Default

Operator: Service tecan

Date:

12-Mar-2014

Signature:



Devices

LiHa Arm

Serial Number: 510000060
 Address: C5
 Firmware Version: V1.40-11/2006
 Bootware Version: V1.00-10/2000

Liquid Channel Configuration

	Tip Type	Pipetting Tubing	Syringe Volume [ul]
Tip 1	Fixed Tip Standard	Standard	1000
Tip 2	Fixed Tip Standard	Standard	1000
Tip 3	Fixed Tip Standard	Standard	1000
Tip 4	Fixed Tip Standard	Standard	1000
Tip 5	Fixed Tip Standard	Standard	1000
Tip 6	Fixed Tip Standard	Standard	1000
Tip 7	Fixed Tip Standard	Standard	1000
Tip 8	Fixed Tip Standard	Standard	1000

Aspiration Tubing Configuration

Tubing type: Standard

FaWa

Serial Number: 1001
 Address: C6T30
 Firmware Version: V2.00-11/99
 Bootware Version: V1.00-11/99

Worktable: Retskemisk fixed tip LD 2014

FaWa Test: Passed

Test Configuration Details

Pump duration before opening the valves[ms]: 1000
 Duration with open valves[ms]: 3000
 Pump duration after closing the valves[ms]: 200
 Tube inner diameter [mm]: 13.0

Pass / Fail Criteria

Minimum expected throughput: see detailed results

Detailed Results

Tips

	Throughput	Min throughput	Within limit
Tip 1, Standard	2030 ul/s	1600 ul/s	Yes
Tip 2, Standard	2039 ul/s	1600 ul/s	Yes
Tip 3, Standard	2083 ul/s	1600 ul/s	Yes
Tip 4, Standard	2101 ul/s	1600 ul/s	Yes
Tip 5, Standard	2052 ul/s	1600 ul/s	Yes
Tip 6, Standard	2021 ul/s	1600 ul/s	Yes
Tip 7, Standard	2044 ul/s	1600 ul/s	Yes
Tip 8, Standard	2061 ul/s	1600 ul/s	Yes

Liquid Level Detection Test: Passed

Test Configuration Details

LLD

Common

Cycles:	15
Clot Error Limit [mm]:	4.0
Error Limit [mm]:	1.5
Air Gap [ul]:	30
LLD Speed [mm/sec]:	60.0
Clot LLD Speed [mm/sec]:	40.0
Source Liquid Conductivity:	Bad
Prefill LLD Mode:	Trough mode
Prefill Aspiration Speed [ul/sec]:	100
Prefill Dispense Speed [ul/sec]:	300
Prefill Submerge [mm]:	2.0
Retract Speed [mm/sec]:	20.0
User prompt on error:	enabled

LLD Source Racks

	Name	FirstWell	WellCount
Trough (Prefill > 5ml)	QCTROUGH.RF	1	1
Microplate	QCTROUGH.RF	1	1
Strip Rack	QCTROUGH.RF	1	1

LLD Dest Racks

	Name	FirstWell	WellCount	Interleaved Wells
Trough (Prefill > 5ml)	QCTROUGH.RM	1	1	0
Microplate	MP96_RoMa.RF	1	8	0
Strip Rack	QCTUBE1.R1	1	8	0

LLD Prefill Volume

	Tip1 [ul]	Tip2 [ul]	Tip3 [ul]	Tip4 [ul]	Tip5 [ul]	Tip6 [ul]	Tip7 [ul]	Tip8 [ul]
Trough (Prefill > 5ml)	0	0	0	0	0	0	0	0
Microplate	200	200	200	200	200	200	200	200
Strip Rack	600	600	600	600	600	600	600	600

LLD Liquid

	Submerge [mm]	Liquid Conductivity	LLD Mode
Trough (Prefill > 5ml)	2.0	Good	Trough mode
Microplate	2.0	Bad	Odd / even tips twice
Strip Rack	2.0	Good	Odd / even tips twice

Pass / Fail Criteria

'z-in-dev' smaller or equal 'Error Limit [mm]': 1.5

'Liquid det err' equals 0

'Clot error' equals 0

How these results are achieved:

For tips that are expected to find liquid (prefill volume > 0 or destination is a trough):

- 1) 'Liquid det err' is incremented if no liquid is detected.
- 2) 'z-in-dev': max difference of the found levels measured over 'Cycles'.
- 3) 'Clot error' is incremented if no exit signal occurs within 'Clot Error Limit'.

For tips that are not expected to find liquid (no prefill and destination is not a trough):

- 1) 'Liquid det err' is incremented if liquid is detected.

2) 'Clot error' is incremented if an exit signal occurs within 'Clot Error Limit'.

Detailed Results

Summary

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-dev	0.3	0.1	0.2	0.2	0.1	0.2	0.2	0.2
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Trough (Prefill > 5ml)

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	42.9	43.1	42.8	42.9	43.1	42.9	43.0	43.5
z-in-max	42.9	43.2	43.0	43.0	43.2	43.0	43.0	43.6
z-in-dev	0.0	0.1	0.2	0.1	0.1	0.1	0.0	0.1
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Microplate

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	72.4	72.1	71.7	71.8	72.0	71.6	71.7	72.0
z-in-max	72.7	72.2	71.8	71.9	72.0	71.8	71.8	72.2
z-in-dev	0.3	0.1	0.1	0.1	0.0	0.2	0.1	0.2
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Strip Rack

	Tip 1	Tip 2	Tip 3	Tip 4	Tip 5	Tip 6	Tip 7	Tip 8
z-in-min	16.8	17.2	17.0	16.5	17.0	17.0	16.6	17.2
z-in-max	16.8	17.2	17.0	16.7	17.0	17.2	16.8	17.2
z-in-dev	0.0	0.0	0.0	0.2	0.0	0.2	0.2	0.0
z-out-min	na	na	na	na	na	na	na	na
z-out-max	na	na	na	na	na	na	na	na
z-out-dev	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Liquid det err	0	0	0	0	0	0	0	0
Clot error	0	0	0	0	0	0	0	0
Passed	yes	yes	yes	yes	yes	yes	yes	yes

Gravimetric Pipetting Precision Test: Not Done

Test Configuration Details

Pass / Fail Criteria

History

Detailed Results

Colorimetric Pipetting Precision Test: Not Applicable

Test Configuration Details

Pass / Fail Criteria

Detailed Results

Kalibreringscertifikat / Calibration Certificate

Nummer/number: 9.12-1820
Task/file: 113-30947
Dato/date: 2013-10-03
Side/page: 1 af 4/1 of 4
Bilag/appendices: Ingen/none.

Opgave/task: Akkrediteret kalibrering af konventionel værdi af masse.
Accredited calibration of conventional value of mass.

Kundeudstyr/items: 1 stk. 10 g E2, 1 stk. 1 g E2, 1 stk. 100 mg E2, 1 stk. 10 mg E2.

Kunde/customer: Tecan Nordic AB Danmark
Himmellevej 29
4000 Roskilde

Attentionperson/for the attention of: Niels Andersen
Ordre/order:

Resultater/results: Se side 2/calibration results are on page 2.

Kal. Dato/date of calibration: 2013-10-03 / 03/10/2013

Kal. af/operator: KLP


Metode/procedure: Instruktionerne 2-09-0206 og 2-09-0208 samt OIML D 28 (2004).
Instructions numbered 2-09-0206 and 2-09-0208. OIML D 28 (2004).

Udstyr/equipment: Se side 4/Equipment is listed on page 4.

Sporbarhed/traceability: Kalibreringsresultaterne er sporbare til den internationale kilogram prototype i Paris. Dansk Fundamental Metrologi A/S vedligeholder sporbarheden.
The mass standards are traceable to the international prototype of the kilogram. The traceability is maintained by Danish Fundamental Metrology A/S.

Underskrifter/signatures:


Flemming Grud Madsen
Kontrollant/Controller
fgm@force.dk


Klaus Poulsen
Underskriftberettiget/Signatory
klp@force.dk

FORCE Technology, Måleteknik, Navervej 1, 6600 Vejen, e-brev: force@force.dk webadresse: www.force.dk
FORCE Technology, Metrology, Navervej 1, 6600 Vejen, Denmark, email: force@force.dk homepage: www.force.dk

Telefon/phone number: +45 76 96 16 00 Telefax/fax number: +45 75 36 41 55

Certifikatet må kun gengives i uddrag med instituttets skriftlige tilladelse.
The certificate may only be partially copied when FORCE Technology has granted permission.

De "almindelige betingelser" på bagsiden er en integreret del af vor ydelse.
The general conditions on the reverse of the page regulate this service.

Udstyrs- betegnelse	Lod- mærkning	Konventionel værdi af masse \pm ekspanderet usikkerhed	Massefylde \pm ekspanderet usikkerhed, dækningsfaktor er 2	Masse: Dæknings- faktor / Friheds- grader Mass: Coverage factor / degrees of freedom	Lodkode
Designation of equipment	Mark on weight	Conventionel value of mass \pm expanded uncertainty	Density \pm expanded uncertainty, coverage factor is 2		Weight Code

Måleværdi tillagt måleusikkerhed er indenfor tolerance: Overensstemmelse.

The measurement result is within the tolerance when added the measurement uncertainty: Compliance.

ZO 158	10,000011 \pm 0,000020 g	7950 \pm 140 kg/m ³	2 / > 50	06-04-02-01-01-10
128235	1,000005 \pm 0,000010 g	7950 \pm 140 kg/m ³	2 / > 50	06-02-02-01-03-10
15836	100,0033 \pm 0,0050 mg	7950 \pm 140 kg/m ³	2 / > 50	05-02-02-01-03-10
15833	10,0022 \pm 0,0026 mg	7950 \pm 140 kg/m ³	2 / > 50	05-02-02-01-03-10

Der er ved beregning af måleusikkerhed anvendt en dækningssandsynlighed på 95 %.
A confidence level of 95 % is used for calculation of measurement uncertainty.

Lodkoden er sammensat af Form, Fabrikat, Materiale, Overflade, Emballage og Klasse
Den har formen Form- Fabrikat-Materiale-Overflade-Emballage-Klasse
Weight Code is composed of Shape, Manufacturer, Material, Surface, Packing and Class.
It is written as Shape-Manufacturer-Material-Surface-Packing-Class.

Form/Shape

- 01 Anden form
Undefined
- 02 Cylinder
Cylindrical
- 03 Blok
Parallelepiped
- 04 Plade
Sheet
- 05 Tråd
Wire
- 06 Cylinder med fladt håndtag
Cylindrical with lifting knob

Form/Shape

- 07 Cylinder med knap
Cylindrical with small knob
- 08 Kroglod
With hook
- 09 Tønde
Barrel shaped
- 10 Ring
Ring shaped
- 11 Omvendt keglestub
Truncated cone
- 12 Kegel med hank
Conical with handle

Form/Shape

- 13 Kegel med knap
Conical with small knob
- 14 Sekskant
Hexagonal
- 15 Kasse
Box shaped
- 16 Skivelod
Disc shaped

Fabrikat/Manufacturer

- 01 Ukendt/Unknown
- 02 Mettler-Toledo
- 03 Haigis
- 04 Zwiebel
- 05 Kern
- 06 Sartorius

Fabrikat/Manufacturer

- 07 Oertling
- 08 Avery
- 09 Ohaus
- 10 Precisa
- 11 SFK
- 12 Crisplant

Fabrikat/Manufacturer

- 13 Novo Nordisk
- 14 Justervæsenet
- 15 Dantest
- 16 FORCE
- 17 Andet/another

Materiale/Material

- 01 Andet
Undefined
- 02 Rustfrit stål
Stainless steel
- 03 Messing
Brass
- 04 Nysølv
Nickel silver

Materiale/Material

- 05 Aluminium
Aluminium
- 06 Gråt støbejern
Cast iron (grey)
- 07 Hvidt støbejern
Cast iron (white)
- 08 Kulstofstål
Carbon steel

Materiale/Material

- 09 Jern
Iron
- 10 Glas
Glass
- 11 Betonkerne
Core of concrete
- 12 Blykerne
Core of lead

Overflade/Surface

- 01 Som materialet
Same as material
- 02 Rustfrit stål
Stainless steel
- 03 Fornikling
Nickel plating
- 04 Forkroming
Chromium plating
- 05 Galvanisering
Galvanisation
- 06 Metallisering
Metal coating
- 07 Varmforzinkning
Hot galvanisation
- 08 Maling
Paint
- 09 Fosfatering
Phosphate coating

Emballage/Packing

- 01 Ingen
None
- 02 Andet
Undefined
- 03 Æske
Box
- 04 Kasse
Box
- 05 Kuffert
Case
- 06 Plastpose
Plastic bag

Klasse/Class

- 01 Ingen/None
- 02 Midterste toleranceklasse
Medium accuracy
- 03 OMIL R111 M3
- 04 OMIL R111 M2-3
- 05 OMIL R111 M2
- 06 OMIL R111 M1-2
- 07 OMIL R111 M1
- 08 OMIL R111 F2
- 09 OMIL R111 F1
- 10 OMIL R111 E2
- 11 OMIL R111 E1
- 12 OMIL R47 n= 1000
- 13 OMIL R47 n= 3000
- 14 OMIL R47 n= 5000
- 15 OMIL R47 n= 10000
- 16 Kundevalgt/specified
- 17 Anden/another

Metode:

FORCE instruktion 2-09-0206 og 2-09-0208 samt OIML D 28 (2004).

Testloddets masse bestemmes ved sammenlignende vejning med massenormaler.
Testloddet vejes eventuelt sammen med en eller flere massenormaler.

Luftens massefylde beregnes i henhold til "Equation for the Determination of the Density of Moist Air /1981/91)".

Måleusikkerheden beregnes som foreskrevet i EA-4/02.

Den rapporterede ekspanderede usikkerhed er angivet som standardusikkerheden multipliceret med en dækningsfaktor som svarer til en dækningssandsynlighed på ca. 95 %. Antallet af effektive frihedsgrader er beregnet og rundet ned til nærmeste heltal. Dækningsfaktoren er bestemt ved opslag i tabel med valgte frihedsgrader f og t -værdier som stammer fra tabel med fraktiler $t_{95,45}$ % i t -fordelingen $t(f)$.

Procedure:

FORCE instruction 2-09-0206 and 2-09-0208 and OIML D 28 (2004).

Calibration is performed by comparing the test weight with one or more mass standards on a mass comparator. The test weight might be weighed together with one or more mass standards.

The airdensity is calculated according to "Equation for the Determination of the Density of Moist Air (1981/91)".

The uncertainty of measurement is calculated according to EA-4/02.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by a coverage factor k . The coverage factor k corresponds for a t -distribution with ν effective degrees of freedom to a coverage probability of approximately 95%.

Udstyr/equipment:

Miljøudstyr Climate measurement equipment	Massekomparator Mass comparator	Massenormaler Mass standards
A70143 x	V01008	X00011 (E2)
A70146 x	V01009	X00012 (E1)
A70147 x	V01010	X00013 (E1) x
A70148	V01016 x	X01011 (F1)
A70149	V01017 x	X01012 (F1)
A70180	V01019	X02002 (F2)
A70181	V01031	X02003 (F2)
A70168 x		X02009 (F2)
A75017 x		Version: 1.00
A75024		Id.: A268-93

Service rapport nr.: 10789342

Rapport udskrevet 01-10-2013

Rekvirent

Tecan Nordic
Himmerlevvej 29
4000 Roskilde

Afdeling
Kontakt Niels Andersen
Kontakt init

Emne (UUT)

Vægt/platform	Fabrikat	Mettler Toledo	Type	AG285
	Serie nr.	1120503480	ID. nr.	
	Kapacitet	210 g		
Placering	Bygning		Rum	

Vejeområde(r)	Fra	Til	Deling
	0,00000 g	80,00000 g	10 ug
	80,0000 g	210,0000 g	100 ug

Betingelser

Rummiljø	Ok	Nivellering	Ok	Vegebord	OK	Rengøring	Ok
Int. lod	Bruges	Auto. nulindtræk	On	Metode	Laboratorie		

Serviceaktiviteter

Følsomhed justeret	Ja	Internt lod		Linearitet justeret	Nej
Nivelleret	Nej	Exc. belastning justeret	Nej	Repareret	Nej
Opstilling	Stabilt arbejdsbord				
Bemærkninger	Vægten meget ustabil grundet manglene hus. Vægten kan ikke akkrediteres.				

Anvendte normaler

1	<u>ID nr.</u> 23725	<u>Certifikat nr.</u> MTm3P00008-K04	<u>Kalibreret dato</u> 23-01-2013	<u>OIML, klasse</u> E2
---	-------------------------------	--	---	----------------------------------

Sporbarhed

Anvendte lodnormaler er sporbare til BIPM via følgende kalibreringslaboratorier:
Mettler Toledo AG i Schweiz (Reg nr 032), SP i Sverige (RMP01) , FORCE (Reg. nr 9) og TI (Reg. nr 200).

Service rapport afsluttet af Rupert Rasmussen 01-10-2013

Dato: 1-10-2013 Navn: R. Rasmussen

Service rapport nr.: 10789342

Rapport udskrevet 01-10-2013

Før service

Repeterbarhed

Gentagelse 01	100,0006 g	Gentagelse 02	100,0009 g
Gentagelse 03	100,0009 g	Gentagelse 04	100,0010 g
Gentagelse 05	100,0009 g		

Anvendt belastning : 100,0001 g

Gennemsnit: 100,00086 g

Minimum: 100,0006 g

Standard afv. 0,00015 g

Maksimum 100,0010 g

S(rel) 0,000152 %

Difference 0,00040 g

Excentrisk afprøvning



Pos 1 start	100,0005 g	Pos 4	100,0009 g
Pos 2	100,0000 g	Pos 5	100,0000 g
Pos 3	100,0004 g	Pos 1 slut	100,0005 g

Anvendt belastning : 100,0001 g

Excentrisk belastningsfejl 0,0005 g

Vejeprobe Laboratorie

Prøve nr.	Nominel belastning	Anvendt belastning	Visningsfejl	Visninger:
<u>1</u>	0,01000 g	0,0100013 g	-0,00009 g	0,00991 g
<u>2</u>	0,10000 g	0,1000025 g	0,00018 g	0,10018 g
<u>3</u>	1,00000 g	0,9999819 g	-0,00003 g	0,99995 g
<u>4</u>	10,00000 g	10,000015 g	0,00023 g	10,00024 g
<u>5</u>	50,00000 g	50,000044 g	0,00010 g	50,00014 g
<u>6</u>	100,0000 g	100,000087 g	0,0003 g	100,0004 g
<u>7</u>	150,0000 g	150,000131 g	0,0005 g	150,0006 g
<u>8</u>	200,0000 g	200,00013 g	0,0008 g	200,0009 g

Service rapport nr.: 10789342

Rapport udskrevet 01-10-2013

Efter service

Repeterbarhed

Gentagelse 01	100,0001 g	Gentagelse 02	100,0000 g
Gentagelse 03	100,0000 g	Gentagelse 04	99,9999 g
Gentagelse 05	99,9998 g		

Anvendt belastning : 100,0001 g

Gennemsnit: 99,99996 g

Minimum: 99,9998 g

Standard afv. 0,00011 g

Maksimum 100,0001 g

S(rel) 0,000114 %

Difference 0,00030 g

Excentrisk afprøvning



Pos 1 start	100,0000 g	Pos 4	100,0000 g
Pos 2	99,9997 g	Pos 5	99,9998 g
Pos 3	100,0000 g	Pos 1 slut	99,9998 g

Anvendt belastning : 100,0001 g

Excentrisk belastningsfejl 0,0003 g

Vejeprøve Laboratorie

Prøve nr.	Nominel belastning	Anvendt belastning	Visningsfejl	Visninger:
<u>1</u>	0,01000 g	0,0100013 g	0,00004 g	
<u>2</u>	0,10000 g	0,1000025 g	-0,00002 g	0,01004 g
<u>3</u>	1,00000 g	0,9999819 g	-0,00001 g	0,09998 g
<u>4</u>	10,00000 g	10,000015 g	-0,00001 g	0,99997 g
<u>5</u>	50,00000 g	50,000044 g	-0,00014 g	10,00000 g
<u>6</u>	100,0000 g	100,000087 g	-0,0001 g	49,99990 g
<u>7</u>	150,0000 g	150,000131 g	-0,0001 g	100,0000 g
<u>8</u>	200,0000 g	200,00013 g	0,0002 g	150,0000 g
				200,0003 g